

LogTag Recorders

New Zealand Office :LogTag Recorders Limited, PO Box 362-95, Northcote, Auckland 0748, New Zealand

CERTIFICATE OF CALIBRATION of

Primary Temperature Reference

The Primary Temperature Reference instrument is used by LogTag Recorders Ltd to calibrate and verify all other in-house temperature reference instruments used for the verification and calibration of LogTag Recorders products. The measurements produced by the primary temperature reference are traceable via this independent calibration to the International System of Metric Units which is in equivalence to NIST standards.

Equipment Details : Michell Instruments Optidew serial # 120442
with PRT probe serial # 120442

Last Calibrated : 28th July 2011

Next Calibration due : 28th July 2012

Calibration by : Measurements Standards Laboratory of New Zealand

Calibration report ref : Temperature/2011/167 (available on request)

Laboratory accreditation : ISO/IEC17025

Calibration table

| Thermometer Reading (°C) | Correction (°C) |
|--------------------------|-----------------|
| -25.0 | +0.0 |
| -19.7 | +0.0 |
| -14.3 | +0.0 |
| -9.0 | +0.0 |
| -3.7 | +0.1 |
| +1.5 | +0.1 |
| +6.8 | +0.1 |
| +12.1 | +0.1 |
| +17.5 | +0.1 |
| +22.7 | +0.1 |
| +28.0 | +0.1 |
| +33.3 | +0.1 |
| +38.7 | +0.1 |
| +44.0 | +0.1 |
| +49.3 | +0.1 |
| +54.6 | +0.1 |
| +59.9 | +0.1 |

ice point reading : 0.0°C

This certificate is issued by : LOGTAG RECORDERS LIMITED

Dated 19th August 2011

Signed :


LogTag Recorders Limited